

HSCC (India) Ltd

AMENDMENT – IX

Dated: - 22.12.2016

Subject: Amendment to the tender Enquiry Document.

Ref: Tender Enquiry No.: HSCC/PUR/AIIMS/RAEBARELI/OPD/2016 Dated 04.10.2016

Bidders are requested to note the following technical Amendment:

Item No.7 Slit lamp

Tender Specification	Amended as
Should have illumination form top using tungsten filament bulb capable of giving Illumination intensity up to 600,00Lux	Instrument to have illumination from LED/Halogen/Tungsten bulb/or equivalent
Slit length 0.2 to 8mm	Slit length 0.2 to 8mm or more
Slit length 0.2 to 8mm	Slit width 0.2 to 8mm or more
Eyepieces of 12.5X	Eyepieces of 10x/12.5x/16x
Should have the facility for stereoscopic examination of fundus where in Angle on stereoscopic observation can be reduced from 13 Deg. To 4.5 Deg. To have better view of eyes with small pupil or high myopia	Delete

Item No. 8 Biochemistry Auto Analyzer

Tender Specification	Amended as
28. The equipment to be supplied should have FDA and CE certification	28. The equipment to be supplied should have FDA or CE certification

Bid sale, submission and opening date for all 2,7,8,9,12 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	29.12.2016, 3.00 P.M
ii.	Closing date & time for receipt of tender	29.12.2016, 3.30 P.M.
iii.	Time and date of Opening of Tenders	29.12.2016, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

Director (PMSSY)
Ministry of Health and Family Welfare,
New Delhi