

Amendment -1 dated 10.11.2015

IFB: HSCC/KCGMC/Medical Equipment /2015-03 dated 22.10.2015

Department of Medicine			
Item no.	Bidders queries /request	Tender Specification	Decision on the Bidder quires
01-ECG Machine – 12 Channels with Interpretation	1.*State of the art 12 channel digital ECG machine with simultaneous acquisition of all leads and an LCD display of the trace prior to printing. It should be Touch Screen. To be added- Equipment should have US FDA certificate.	1*State of the art 12 channel digital ECG machine with simultaneous acquisition of all leads and an LCD display of the trace prior to printing	As per tender
02- Video Colonoscope Complete Set	1. *Should be compatible with leak testing facility.	1. *Should be compatible with semi automatic leakage tester	As per tender
	3. Outputs - RGB, Y/C, VBS Composite	3.3. Outputs - RGB, Y/C, VBS Composite, XGA & DV/ simultaneous	As per tender
	4. It should have structure/surface/edge enhancement option for better image quality	4. It should have structure and edge enhancement option for better image quality	“ It should have structure and edge enhancement option for better image quality” may please read as “ It should have structure/surface/edge enhancement option for better image quality “
	7. Light source - Combined or separate 150 watts or more Halogen/100 watt or more Xenon Light Source	7*Light source - Combined or separate 150 watts or more Xenon Light Source	As per tender
	Specifications and Qualifications: Angle of view field:-100 ° or more 140 deg, 100 deg will be too less	Specifications and Qualifications: Angle of view field:-140 ° or more	“Angle of view field:-140 ° or more” may please read as “Angle of View 100 degree or more”
	Depth of View: -3-100mm	Depth of View: 4-100mm or more	“Depth of View: 4-100mm or more” may please read as ‘Depth of view 3-100mm or more”
	Diameter of Instrument Channel minimum 3.8mm	Diameter of Instrument Channel:-ø3.2mm or more	As per tender
	15” is too small; it should be increased to 19”	Medical Grade Color Monitor: 15" or more Medical grade color monitor	As per tender

Amendment -1 dated 10.11.2015

IFB: HSCC/KCGMC/Medical Equipment /2015-03 dated 22.10.2015

Department of Paediatric

Item no.	Bidders queries /request	Tender Specification	Decision on the Bidder quires
Item no.2- Infant Incubator	M/s Vishal Surgical Equipment 3.1 (2)* Two shelves cabinet with door/without door.	3.1 (2)* Two shelves cabinet with door.	"Two shelves cabinet with door" may please read as Two shelves cabinet with door/without door.