

Amendment-XXIV**Ref.: IFB No. HSCC/SJH/Medical Equipment/2015 Dated 10.4.2015**

Sub.: Procurement of Medical Equipment for New Emergency Block & Super-Specialty Block at Safderjung Hospital, New Delhi.

Following amendments are issued:**S.NO. 2 Central Monitoring Stations (CMS) 2 nos. with bedside Monitors 12 nos. for Heart Command Centre for Superspeciality Block**

S. No.	Tender Specifications	Amended Specifications
1	Point 3.2 Should have battery back-up for one hour or more.	Deleted
2	Point 3.7 Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- channel IBP with thermal recorder. NMT integrated / standalone. Upgradable to monitor cardiac output	Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- channel IBP with thermal recorder. NMT integrated Upgradable to monitor cardiac output
3	Point 3.9 EtCO2 – side stream. Display both inspired and expired values, showing capnography.	EtCO2 – side stream/ main stream / micro stream Display both inspired and expired values, showing capnography.

Amendment to be issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged.

**Medical Superintendent
Safderjung Hospital & VMMC,
New Delhi.**