

Amendment-XXIX**Ref.: IFB No. HSCC/SJH/Medical Equipment/2015 Dated 10.4.2015**

Sub.: Procurement of Medical Equipment for New Emergency Block & Super-Specialty Block at Safderjung Hospital, New Delhi.

As per DGHS letter no. Z-28015/53/2015-MH-1 dt. 22.9.2015 (enclosed) following amendments to be issued for Item No. 1, 2, 3, 4 & 9:

S.NO. 1 Central Monitoring Stations (CMS) 8 nos. with bedside Monitors 125 nos. for ICU for Super speciality Block

S.No.	Tender Specifications	Amended Specification
1	Point 3.7 Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2-Channel IBP with thermal recorder. NMT intergraded/ standalone. Upgradable to monitor cardiac output.	Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- Channel IBP with thermal recorder. Facility of intergraded NMT Upgradable to monitor cardiac output.

S.NO. 2 Central Monitoring Stations (CMS) 2 nos. with bedside Monitors 12 nos. for Heart Command Centre for Super speciality Block

S.No.	Tender Specifications	Amended Specification
1	Point 3.7 Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2-Channel IBP with thermal recorder. NMT intergraded/ standalone. Upgradable to monitor cardiac output.	Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- Channel IBP with thermal recorder. Facility of intergraded NMT Upgradable to monitor cardiac output.

S.NO. 3 Central Monitoring Stations (CMS) 2 nos. with bedside Monitors 20 nos. for Cardiology for Super-Specialty Block

S.No.	Tender Specifications	Amended Specification
-------	-----------------------	-----------------------

1	Point 3.7 Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- Channel IBP with thermal recorder. NMT intergraded/ standalone. Upgradable to monitor cardiac output.	Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- Channel IBP with thermal recorder. Facility of intergraded NMT Upgradable to monitor cardiac output.
----------	--	--

S.NO. 4 Central Monitoring Stations (CMS) 5 nos. with bedside Monitors 68 nos. for ICU for Emergency Block

S.No.	Tender Specifications	Amended Specification
1	Point 3.7 Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- Channel IBP with thermal recorder. NMT intergraded/ standalone. Upgradable to monitor cardiac output.	Facility to monitor and display –ECG, Respiration, NIBP, SpO2- Massimo technology, CO2 with capnography, two Temperatures & 2- Channel IBP with thermal recorder. Facility of intergraded NMT Upgradable to monitor cardiac output.

S.NO. 9 Multipara Monitor (RESP., TEMP., Pulse, SpO2) for Examination Resuscitation & Wards

S.No.	Tender Specifications	Amended Specification
1	Point 3.3 Monitoring parameters:- ECG (3 and 5 lead), respiration, NIBP, Massimo Spo2, temperature IBP Dual and ETCO2 (Side stream)	Monitoring parameters:- ECG (3 and 5 lead), respiration, NIBP, Spo2, temperature IBP Dual and ETCO2 (Side stream)
2	Point 3.6 Trend of at least 72 hours.	Trend of at least 24 hours .
3	Point 4.3 Adult, Pediatric & Neonatal Cuff – 02 each	Adult, NIBP Cuffs 2 per monitor = 812 Pediatric NIBP Cuffs =100 Neonatal NIBP Cuffs =50 Adult SpO2 Probe (2 per monitor) = 812 Pediatric SpO2 Probe = 100 Neonatal SpO2 Probe = 50
4	Point 4.6 Modular/Standalone recorder -01	TO be deleted

For Item No. 9: Quantity Changed from 406 to 370.

Revised EMD: Rs. 21,18,200.00

Bid submission date remains unchanged i.e. 08.10.2015.

Amendment issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged.

**Medical Superintendent
Safderjung Hospital & VMMC,
New Delhi.**