

All Bidders

Amendment-2

Subject: Tender for Package-II: Supply, installation, testing & commissioning of Hospital & office Furniture & other associated works like minor civil, electrical, PHE works etc. for additional floors of Trauma Centre in Trauma Centre of AIIMS, Ansari Nagar, New Delhi

Dear Sir,

This has reference to the pre bid meeting held on 03.05.2016 for captioned subject works,

Please find enclosed the revised specifications of few items at Annexure-1 as an AMMENDMENT-2/CLARIFICATION-2 may be noted, which shall be treated as a part of subject tender.

The following Amendments also may be noted as under:

Sl no	Existing Provision	Amended provision
1	As per Prequalification Clause no 3.2 (iii) : Similar works means: Supply & Installation of Hospital/Health Care furniture works. Bidders must have executed at least one work of office furniture of any value.	May be read as : As per Prequalification Clause no 3.2 (iii) : Similar works means: Supply & Installation of Hospital/Health Care furniture works.

BOQ item no	Specification as per Amendment-1	Amended specification
19	Trauma Care Trolley- Should be a portable screen on castor having three screens which should be folded to provide privacy in ward. Overall size 2500-2650 mm (L) X550-640 mm (W) X1650-1750 mm (H). Fix frame of ERW tube of section 25.4 mm dia and thickness 1.2 mm .movable frame of ERW tube 19.05 mm dia and 1.2 mm thick. Flexible frame should have plastic hinge which can be adjusted in provided space. Leg frame should have ERW tube of sq 30 sections and 1.6 mm thickness. 50 mm castor	Trauma Care Trolley may be read as - Stainless steel finished manual height adjustable trolley used in patient mobility for trauma patient during emergency in hospitals. Dimensions (L) 1880-1900 x (W) 710-760mm x (H) upto 910 mm .Material of Construction: Base frame, top frame, mattress support platform and linkage mechanism are made of stainless steel Grade 304. 1) Mechanism structure round tube thickness is Min. 3mm. 2) Square/Rectangular tube thickness is min. 2 mm. 3) Plate used for load bearing members at mechanism has min. thickness 4mm. 4) Plastic fittings are in Nylon. Backrest should be adjusted with the help of crank handle from 0 to 80°,Kneerest can be

	<p>wheel provided for ease of movement. Plastic bush for folding pipe .screen should be of casement fabric for screen which is opaque for providing modesty. powder coating of epoxy polyestar finish</p>	<p>adjusted with the help of crank handle from 0 to 40° Height should be adjusted with the help of crank handle. Retractable type crank handle with foldable lever provides the benefit to user during operation and can be slide in when not in use and thus occupying less space Rubber extruded profile provided around the top frame periphery which provides the safety to stretcher trolley and wall as well from all the direction. Side rail provided at either side for the safety. It should make of aluminum with anodizing finish for better look, feel and enhanced aesthetic. 150 mm wheel dia castor provided with lock in 3 castors and direction lock provided in one castor which provides the safety and stability when not in use. Directional lock castor makes the handling easy during the use. Castor make should have Load carrying capacity per wheel is 120Kg. Inbuilt and foldable type IV poles provided both side for the ease of application Long space at base has been provided with SS 304 tray which provide the facility to user to keep additional accessories into it. At base frame Oxygen cylinder case provided to accommodate the oxygen cylinder for the emergency conditions. Mattress with PU Foam,PVC leatherette cover provided for the patient comfort. PU foam density is 22-25Kg/m³.Mattress has waterproof, fireretardent, anti-bacterial features.</p>

Amendments to be submitted online duly signed & stamp along with tender. Other conditions of tender will remain unchanged.

Prospective bidders are advised to regularly scan through HSCC e-tender portal <http://www.tenderwizard.com/HSCC> and <http://www.hsccltd.co.in> as corrigendum/amendments etc., if any, will be notified on these websites only and separate advertisement will not be made for this

Thanking you,

Project Coordinator, HSCC (I) Ltd