

Amendment-XXX**Ref.: IFB No. HSCC/SJH/Medical Equipment/2015 Dated 10.4.2015**

Sub.: Procurement of Medical Equipment for New Emergency Block & Super-Specialty Block at Safderjung Hospital, New Delhi.

As per DGHS letter no. Z-28015/53/2015-MH-1 dt. 01.10.2015 (enclosed) following amendments to be issued for **Item No. 9 (Multiparameter Monitors)**:

S.NO. 9 Multipara Monitor (RESP., TEMP., Pulse, SpO2) for Examination Resuscitation & Wards

S.No.	Tender Specifications	Amended Specification
1	Point 3.3 Monitoring parameters:- ECG (3 and 5 lead), respiration, NIBP, Spo2, temperature, IBP Dual and ETCO2 (Side stream)	Monitoring parameters:- ECG (3 and 5 lead), respiration, NIBP, Spo2, temperature, IBP Dual and ETCO2 (Side stream/main stream/micro stream)

Amendment issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged.

**Medical Superintendent
Safderjung Hospital & VMMC,
New Delhi.**