

Amendment-VII**Ref.: IFB No. HSCC/SJH/Medical Equipment/2015 Dated 10.04.2015**

Sub.: Procurement of Medical Equipment for New Emergency Block & Super-Specialty Block at Safderjung Hospital, New Delhi.

Following amendment is issued:

To be read as:**Item No. 5 : ICU Ventilator:**

Sr. No.	Specification Lines as per Tender	Amended changes
8	Point 3.a Tidal volume: minimum 50 ml-1500ml or more in volume control. Tilting facility for flexibility	Tidal volume: minimum 50 ml-1500ml or more in volume control.
12	Point 7 Ventilator should have synchronized in built ultrasound nebulizer should give less than or equal to 3 micron particle size	Ventilator should have synchronized inbuilt / ultrasound nebulizer should give less than or equal to 3 micron particle size.
	Point 8 Complied trend analysis at least for 72 hours for all measured parameters. Automatic patient detection	Complied trend analysis at least for 72 hours for all measured parameters

Amendment to be issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged.

**Medical Superintendent
Safderjung Hospital & VMMC,
New Delhi.**