

AMENDMENT No – VI Dated 14.10.2014

Sub.: Procurement of Medical Equipment for NEIAH, Shillong.

Ref: Tender Enquiry No.: HSCC/PUR/NEIAH/Medical Equipment/2014 dated 19.8.2014

ITEM NO 12: VIDEO COLONOSCOPY SYSTEM

Existing Tender Specification	Amended/Read As
Light source - Combined or separate 100 watt or more Xenon with emergency backup facility.	Light source - Combined or separate 150 watts or more Halogen/ 100 watt or more Xenon Light Source.

Bidders should submit all the documents as mentioned in the Check List, GIT Clause 11, Section-IX and elsewhere mentioned in the Bid Document.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Director, NEIAH Shillong